

CERTIFICATE

SAMPLER | VIALS | TEFLON WARE
HYDROTHERMAL | SS CLEAN ROOM ITEMS



MATERIAL SPECIFICATION CERTIFICATE

Part Number : 31316

MSCNo. SAMPLE

Part Name : INERT PRESSURE REACTOR



Diameter : 75 mm

Liner Dia : 49 mm

Height : 154 mm

Liner Height : 93 mm

Liner Moc : Boro Glass 3.3

Body Moc: SS 316L

Max Temp : 380°C

Max PSI : 56.8



JN SCIENCETECH

Works : # 7-4-232, Geetha Nagar, Ferozguda,
Balanagar, Hyderabad - 500 011.

✉ info@jnsciencetech.com | info@jnsciencetech.in

🌐 www.jnsciencetech.com | www.jnsciencetech.in

ISO 9001:2015 CERTIFIED