

# CERTIFICATE

SAMPLER | VIALS | TEFLON WARE  
HYDROTHERMAL | SS CLEAN ROOM ITEMS



## MATERIAL SPECIFICATION CERTIFICATE

Part Number : 13381

MSC No. : sample

Part Name : Oxygen Combustion Flask

capacity : 500ml

Material : Boro 3.3 glass (Heavy Wall)

Sample Holder : Platinum

Platinum Purity : 99.95%

Sample Holder Model : Basket Type

Sample Holder Length : 60 mm

Flask Height : 175 mm

Flask Dia : 107 mm dia

Neck Size : B-29/32



Note: Platinum Sample Holder are two type

1. Flat Sample Holder
2. Basket Sample Holder



### JN SCIENCETECH

Works : # 7-4-232, Geetha Nagar, Ferozguda,  
Balanagar, Hyderabad - 500 011.

✉ info@jnsciencetech.com | info@jnsciencetech.in

🌐 www.jnsciencetech.com | www.jnsciencetech.in

ISO 9001:2015 CERTIFIED